



Plasminogen Deficient Plasma

Product No. ADG240

American Diagnostica GmbH

Kaplaneigasse 35 • D-64319 Pfungstadt • Germany
Tel: +49 (0)6157 990899 • Fax: +49 (0) 6157 990808

Preparation

Plasminogen Deficient Plasma is manufactured from pooled normal human plasma depleted of Plasminogen using antibodies directed to Plasminogen. Only the highest quality citrated plasma is used as starting material.

Warning

The plasma is prepared from screened normal citrated human plasma that have been tested and found negative for HIV and Hepatitis by FDA-approved methods.

As no known test method can provide complete assurance that products derived from human blood will not transmit HBsAg, HCV, HIV-1, HIV-2 or other blood-borne pathogens, this reagent should be handled as recommended for any potentially infectious human specimen.

Application

For research assays of plasminogen antigen and functionality, as a negative control and for dilution of standards and test samples.

Presentation

Clear screw-capped vial of 1 mL of lyophilized plasma. The plasma contains 50mM HEPES and stabilizers.

Reconstitution and Stability

One vial of lyophilized product is reconstituted by addition of 1 mL of distilled water, employing gentle agitation for 15 minutes, at room temperature. Stability after reconstitution is 4 hours at ambient (18-25°C), or 30 days at -20 °C.

Storage

Prior to reconstitution, dried plasma should be stored at 2-8°C.

ADG240©ADG30032012