



Labor Diagnostika Nord GmbH & Co. KG

Am Eichenhain 1, 48531 Nordhorn

Telefon: +49-5921-8197 0

Telefax: +49-5921-8197 222

e-mail: manz@ldn.de

Internet: <http://www.ldn.de>



Instructions for use

IGF-1 ELISA

REF

ME E-0500



IVD



IGF-1 ELISA

Introduction

The **IGF-1 600 Enzyme Immunoassay Kit** provides materials for the quantitative determination of Insulin-like Growth Factor 1 (IGF-1) in serum.

This assay is intended for in vitro diagnostic use only.

Clinical Significance

Insulin-like growth factor I (IGF-1) is a polypeptide of 70 amino acids (7650 daltons), and is one of a number of related insulin-like growth factors present in the circulation. The molecule shows approximately 50% sequence homology with proinsulin and has a number of biological activities similar to insulin. The peptide is growth hormone (GH) dependent to a high degree, but there is growing evidence of GH-independent secretion. IGF-1 has numerous growth-promoting effects, including mitogenic effects and the promotion of cartilage sulphation. It may also be implicated in the rate of bone turnover.

Almost all (>95%) of serum IGF-1 circulates bound to specific IGF binding proteins, of which six classes (IGF-BPs 1 - 6) are now recognized. BP3 is thought to be the major binding protein of IGF-1, forming a ternary complex of 140 000 daltons with IGF-1 and an acid-labile subunit.

Clinical Applications

The measurement of serum IGF-1 is of recognized value in children with growth disorders and in the diagnosis and monitoring of acromegaly. IGF-1 concentrations change with age, nutritional status, body composition and growth hormone secretion.

A single basal IGF-1 determination is useful in the assessment of short stature in children and in nutritional support studies of acutely ill patients. For the diagnosis of acromegaly, a single IGF-1 determination is considered more reliable than a random GH determination.

PRINCIPLE of the test

The IGF-1 600 ELISA Kit is a solid phase enzyme-linked immunosorbent assay (ELISA), based on the principle of competitive binding.

Patient samples, standards and controls are acidified and neutralized prior to the assay procedure.

The microtiter wells are coated with a monoclonal antibody directed towards an antigenic site on the IGF-1 molecule.

The pre-treated sample is incubated at room temperature with Conjugate (biotinylated IGF-1). The wells are washed and then incubated with Enzyme Complex (Streptavidin-HRP-Complex).

After addition of the substrate solution, the intensity of colour developed is reverse proportional to the concentration of IGF-1 in the patient sample.

Precautions

- This kit is for in vitro diagnostic use only.
- For information on hazardous substances included in the kit please refer to Material Safety Data Sheets.
- All reagents of this test kit which contain human serum or plasma have been tested and confirmed negative for HIV I/II, HBsAg and HCV by FDA approved procedures. All reagents, however, should be treated as potential biohazards in use and for disposal.
- Avoid contact with Stop Solution containing 0.5 M H₂SO₄. Avoid contact with 1M HCl for sample acidification and 1M NaOH for neutralization. It may cause skin irritation and burns.
- Never pipet by mouth and avoid contact of reagents and specimens with skin and mucous membranes.
- Do not smoke, eat, drink or apply cosmetics in areas where specimens or kit reagents are handled.
- Wear disposable latex gloves when handling specimens and reagents. Microbial contamination of reagents or specimens may give false results.
- Handling should be in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
- Do not use reagents beyond expiry date as shown on the kit labels.
- All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes and microtiterplate readers.
- Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
- Chemicals and prepared or used reagents have to be treated as hazardous waste according the national biohazard safety guideline or regulation.
- Safety Data Sheets for this product are available upon request directly from the manufacture.
The Safety Data Sheets fit the demands of: EU-Guideline 91/155 EC.

Kit Components

Contents of the Kit

HCL **ME E-0519** **1 M HCl**

1 vial, 3 mL, ready to use; for sample acidification.

NaOH **ME E-0587** **1 M NaOH**

1 vial, 3 mL, ready to use; for neutralization.

96 **ME E-0531** **Microtiterwells**

12x8 (break apart) strips, 96 wells; Wells coated with a anti-IGF-1 antibody (monoclonal).

Standards

	Cat. no.	Standard	Concentration	Volume/Vial
STANDARD A	ME E-0501	Standard 0	0 ng/ml	1 ml
STANDARD B	ME E-0502	Standard 1	5 ng/ml	1 ml
STANDARD C	ME E-0503	Standard 2	10 ng/ml	1 ml
STANDARD D	ME E-0504	Standard 3	50 ng/ml	1 ml
STANDARD E	ME E-0505	Standard 4	150 ng/ml	1 ml
STANDARD F	ME E-0506	Standard 5	300 ng/ml	1 ml
STANDARD G	ME E-0507	Standard 6	600 ng/mL	1 mL

The standards are calibrated against the International Reference Reagent for IGF-1,

NIBSC code: 87/518

conversion: 1ng/mL x 0.13 = nmol/L

contain 0.03% Proclin 300 + 0.005% gentamicin sulfate as a preservative.

CONJUGATE **ME E-0540** **Enzyme Conjugate**

1 vial, 14 mL, ready to use; biotinylated IGF-1;

* contains 0.01% MIT and 0.02% BND as a preservative.

ENZYME **ME E-0515** **Enzyme Complex**

1 vial, 20 mL, ready to use; Streptavidin HRP Complex.

SUBSTRATE **FR E-0055** **Substrate Solution**

1 vial, 14 mL, ready to use; Tetramethylbenzidine (TMB).

STOP-SOLN **FR E-0080** **Stop Solution**

1 vial, 14 mL, ready to use; contains 0.5M H₂SO₄.

Avoid contact with the stop solution. It may cause skin irritations and burns.

WASH- CONC 40x **FR E-0030** **Wash Solution**

1 vial, 30 mL (40X concentrated); see „Preparation of Reagents“.

* BND = 5-bromo-5-nitro-1,3-dioxane

MIT = 2-methyl-2H-isothiazol-3-one

Note: Additional *Standard 0* for sample dilution is available on request.

Equipment and material required but not provided

- A microtiter plate calibrated reader (450±10 nm).
- Calibrated variable precision micropipettes.
- Absorbent paper.
- Aqua dest.
- 1.5 mL Reaction Cups (e.g. from Eppendorf) for Sample Preparation (Acidification and Neutralization).
- Universal Indicator Paper.

Storage and stability of the Kit

When stored at 2-8°C unopened reagents will retain reactivity until expiration date. Do not use reagents beyond this date.

Opened reagents must be stored at 2-8°C. Microtiter wells must be stored at 2-8°C. Once the foil bag has been opened, care should be taken to close it tightly again. Opened microtiter wells retain activity for 6 weeks if stored as described above.

Preparation of Reagents

Allow all reagents and required number of strips to reach room temperature prior to use.

Wash Solution

Dilute 30 mL of concentrated Wash Solution with 1170 mL deionized water to a final volume of 1200 mL.

The diluted Wash Solution is stable for 2 weeks at room temperature.

Disposal of the Kit

The disposal of the kit must be made according to the national regulations. Special information for this product is given in the Material Safety Data Sheets (see chapter 13).

Damaged Test Kits

In case of any severe damage of the test kit or components, the manufacture have to be informed written, latest one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.

SPECIMEN

Serum can be used in this assay.

NOTE: In plasma significantly reduced values were observed.

Do not use haemolytic, icteric or lipaemic specimens.

Please note: Samples containing sodium azide should not be used in the assay.

Specimen Collection

Serum:

Collect blood by venipuncture (e.g. Sarstedt Monovette # 02.1388.001), allow to clot, and separate serum by centrifugation at room temperature. Do not centrifuge before complete clotting has occurred. Patients receiving anticoagulant therapy may require increased clotting time.

Specimen Storage

Specimens should be assayed immediately.

Specimens held for a longer time (up to two months) should be frozen only once at -20°C prior to assay. Thawed samples should be inverted several times prior to testing.

Specimen Dilution

If in an initial assay, a specimen is found to contain more than the highest standard, the specimens can be diluted with *Standard 0* and reassayed as described in Assay Procedure.

For the calculation of the concentrations this dilution factor has to be taken into account.

Example:

a) Dilution 1:10: 10 µL Serum + 90 µL *Standard 0* (mix thoroughly)

b) Dilution 1:100: 10 µL dilution a) 1:10 + 90 µL *Standard 0* (mix thoroughly).

test procedure

General Remarks

- All reagents and specimens must be allowed to come to room temperature before use. All reagents must be mixed without foaming.
- Once the test has been started, all steps should be completed without interruption.
- Use new disposal plastic pipette tips for each standard, control or sample in order to avoid cross contamination.
- Absorbance is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.

Acidification and Neutralization of Patient Samples and Standards

1. Pipette 200 μ L Sample, Standard or Control in 1.5 mL-Reaction caps (E.g. Eppendorf-Caps).
Please note: The standards should be acidified and neutralized too, according to the procedure described below
2. Add 20 μ L 1 M HCl.
3. Mix and incubate for 15 minutes.
4. For Neutralization add 20 μ L 1M NaOH to all caps and mix the solution.
5. After Neutralization the pH value should be 7 to 8.
Please check the pH value with Universal Indicator Paper.

Assay Procedure

Each run must include a standard curve.

1.	Secure the desired number of Microtiter wells in the holder.
2.	Dispense 50 μL of each acidified and neutralized <i>Standard</i> , controls and samples <u>with new disposable tips</u> into appropriate wells.
3.	Dispense 100 μL <i>Enzyme Conjugate</i> into each well. Thoroughly mix for 10 seconds. It is important to have a complete mixing in this step.
4.	Incubate for 120 minutes at room temperature.
5.	Briskly shake out the contents of the wells. Rinse the wells 3 times with diluted Wash Solution (400 μ L per well). Strike the wells sharply on absorbent paper to remove residual droplets. Important note: The sensitivity and precision of this assay is markedly influenced by the correct performance of the washing procedure!
6.	Dispense 150 μL <i>Enzyme Complex</i> into each well.
7.	Incubate for 60 minutes at room temperature.
8.	Briskly shake out the contents of the wells. Rinse the wells 3 times with diluted Wash Solution (400 μ L per well). Strike the wells sharply on absorbent paper to remove residual droplets.
9.	Add 100 μL of <i>Substrate Solution</i> to each well.
10.	Incubate for 30 minutes at room temperature.
11.	Stop the enzymatic reaction by adding 100 μL of <i>Stop Solution</i> to each well.
12.	Read the OD at 450$\pm$10 nm with a microtiter plate reader within 10 minutes after adding the Stop Solution.

Calculation of Results

1. Calculate the average absorbance values for each set of standards, controls and patient samples.
2. Construct a standard curve by plotting the mean absorbance obtained from each standard against its concentration with absorbance value on the vertical(Y) axis and concentration on the horizontal (X) axis.
3. Using the mean absorbance value for each sample determine the corresponding concentration from the standard curve.
4. Automated method: The results in the IFU have been calculated automatically using a 4 PL (4 Parameter Logistics) curve fit. 4 Parameter Logistics is the preferred method. Other data reduction functions may give slightly different results.
5. The concentration of the samples can be read directly from this standard curve. Samples with concentrations higher than that of the highest standard have to be further diluted. For the calculation of the concentrations this dilution factor has to be taken into account.

Example of Typical Standard Curve

The following data is for demonstration only and cannot be used in place of data generations at the time of assay.

Standard	Optical Units (450 nm)
Standard 0 (0 ng/mL)	1.94
Standard 1 (5 ng/mL)	1.69
Standard 2 (10 ng/mL)	1.49
Standard 3 (50 ng/mL)	0.93
Standard 4 (150 ng/mL)	0.56
Standard 5 (300 ng/mL)	0.44
Standard 6 (600 ng/mL)	0.33

Expected values

It is strongly recommended that each laboratory should determine its own normal and abnormal values.

Children at pre-puberty (age 3-8 years): 20 – 250 ng/mL
Smaller children with a low body weight could show values under 50 ng/mL.

Children at puberty (age 11 – 16 years): 130 – 600 ng/mL
Adults after puberty: 150 – 350 ng/mL

Note: The IGF-1 level could slightly decrease at age.

Quality Control

Good laboratory practice requires that controls be run with each calibration curve. A statistically significant number of controls should be assayed to establish mean values and acceptable ranges to assure proper performance.

It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results. Use controls at both normal and pathological levels. The controls and the corresponding results of the QC-Laboratory are stated in the QC certificate added to the kit. The values and ranges stated on the QC sheet always refer to the current kit lot and should be used for direct comparison of the results.

It is also recommended to make use of national or international Quality Assessment programs in order to ensure the accuracy of the results.

Employ appropriate statistical methods for analysing control values and trends. If the results of the assay do not fit to the established acceptable ranges of control materials patient results should be considered invalid.

In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods.

After checking the above mentioned items without finding any error contact your distributor or the manufacture directly.

Assay Characteristics

Assay Dynamic Range

The range of the assay is between 0 – 600 ng/mL.

Analytical Sensitivity

The analytical sensitivity was calculated from the mean minus two standard deviations of twenty (20) replicate analyses of *Standard 0* and was found to be 1.292 ng/mL.

Precision

Intra Assay Variation

The within assay variability is shown below:

Sample	n	Mean (ng/mL)	CV (%)
1	20	29.04	4.72
2	20	58.94	6.62

Inter Assay Variation

The between assay variability is shown below:

Sample	n	Mean (ng/mL)	CV (%)
1	12	28.06	7.22
2	12	57.52	7.79

Recovery

Samples have been spiked by adding IGF-1 solutions with known concentrations in a 1:1 ratio. The expected values were calculated by addition of half of the values determined for the undiluted samples and half of the values of the known solutions. The % Recovery has been calculated by multiplication of the ratio of the measurements and the expected values with 100.

Sample	Added Concentration 1:1 (v/v) (ng/mL)	Measured Conc. (ng/mL)	Expected Conc. (ng/mL)	Recovery (%)
1	0.00	28.86		
	300.00	294.30	314.43	93.6
	150.00	152.43	164.43	92.7
	75.00	77.04	89.43	86.1
	25.00	38.48	39.43	97.6
2	0.00	56.36		
	300.00	334.80	328.18	102.0
	150.00	191.28	178.18	107.4
	75.00	99.98	103.18	96.9
	25.00	46.77	53.18	87.9

Linearity

Sample	Dilution	Measured Conc. (ng/mL)	Expected Conc. (ng/mL)	Recovery (%)
1	None	28.86	28.86	
	1:2	13.85	14.43	96.0
	1:4	7.31	7.22	101.3
	1:8	3.78	3.61	104.9
	1:16	1.85	1.80	102.6
2	None	56.36	56.36	
	1:2	26.77	28.18	95.0
	1:4	13.29	14.09	94.3
	1:8	6.75	7.05	95.8
	1:16	3.56	3.52	101.1

Limitations of Use

Any improper handling of samples or modification of this test might influence the results.

Interfering Substances

Haemoglobin (up to 2 mg/mL), Bilirubin (up to 0.25 mg/mL) and Triglyceride (up to 30 mg/mL) have no influence on the assay results.

Drug Interferences

Until today no substances (drugs) are known to us, which have an influence to the measurement of IGF-1 in a sample.

High-Dose-Hook Effect

No hook effect was observed in this test.

Legal Aspects

Reliability of Results

The test must be performed exactly as per the manufacturer's instructions for use. Moreover the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable national standards and/or laws. This is especially relevant for the use of control reagents. It is important to always include, within the test procedure, a sufficient number of controls for validating the accuracy and precision of the test.

The test results are valid only if all controls are within the specified ranges and if all other test parameters are also within the given assay specifications. In case of any doubt or concern please contact the manufacture.

Therapeutic Consequences

Therapeutic consequences should never be based on laboratory results alone even if all test results are in agreement with the items as stated under point 11.1. Any laboratory result is only a part of the total clinical picture of a patient.

Only in cases where the laboratory results are in acceptable agreement with the overall clinical picture of the patient should therapeutic consequences be derived.

The test result itself should never be the sole determinant for deriving any therapeutic consequences.

Liability

Any modification of the test kit and/or exchange or mixture of any components of different lots from one test kit to another could negatively affect the intended results and validity of the overall test. Such modification and/or exchanges invalidate any claim for replacement.

Claims submitted due to customer misinterpretation of laboratory results subject to point 11.2. are also invalid. Regardless, in the event of any claim, the manufacturer's liability is not to exceed the value of the test kit. Any damage caused to the test kit during transportation is not subject to the liability of the manufacturer.

REFERENCES

1. Daughaday E, Rotwein P: Insulin like growth factors I and II. Peptide, messenger ribonucleic acid and gene structures, serum and tissue concentrations. *Endocrin Rev* 10:68-91, 1989.
2. Baxter RC, Martin JL, Beniac VA: High molecular weight insulin-like growth factor binding protein complex. *J Biol Chem* 264:11843-11848, 1989.
3. Rechler M: Insulin-like growth factor binding proteins. *Vit Horm* 47:1-114, 1993.
4. Zapf J, Hauri C, Waldvogel M, Froesch ER: Acute metabolic effects and half-lives of intravenously administered insulin-like growth factors I and II in normal and hypophysectomized rats. *J Clin Invest* 77:1768-1775, 1986.
5. Guler HP, Zapf J, Froesch ER: Short-term metabolic effects of recombinant human insulin-like growth factor-I in healthy adults. *New Engl J Med* 317:1237-140, 1987.
6. Costigan DC, Guyda HJ, Posner BI: Free insulin-like growth factor I (IGF-I) and IGF-II in human saliva. *J Clin Endocrinol Metab* 66:1014-1018, 1988.
7. Lewitt MS, Denyer GS, Cooney GJ, Baxter RC: Insulin-like growth factor-binding protein-1 modulates blood glucose levels. *Endocrinology* 129:2254-2256, 1991.
8. Lewitt MS, Saunders H, Baxter RC: Bioavailability of insulin-like growth factors (IGFs) in rats determined by the molecular distribution of human IGF-binding protein-3. *Endocrinology* 133:1797-1802, 1993
9. Lieberman SA, Bukar J, Chen SA, Celniker AC, Compton PG, Cook J, Albu J, Perlman AJ, Hoffman AR: Effects of recombinant human insulin-like growth factor-I (rhIGF-I) on total and free IGF-I concentrations, IGF-binding proteins, and glycemic response in humans. *J Clin Endocrinol Metab* 75:30-36, 1992.
10. Schneiderman R, Maroudas A, Lee PDK: Concentrations of IGF-I and its complexes in normal and osteoarthritic human cartilage: in situ values. *Orthopedic Res Soc*, submitted, 1994.

Symbols:

	Storage temperature		Manufacturer		Contains sufficient for <n> tests
	Expiry date		Batch code		For in-vitro diagnostic use only!
	Consult instructions for use		Content		CE labelled
	Caution		Catalogue number		For research use only!