

Instructions for use

Insulin ELISA

REF

ME E-0900



IVD



Insulin ELISA

Introduction

The **Insulin Enzyme Immunoassay Kit** provides materials for the quantitative determination of Insulin in serum and plasma (Heparin- or Citrate-plasma).

This assay is intended for in vitro diagnostic use only.

Insulin is the principal hormone responsible for the control of glucose metabolism. It is synthesized in the β -cells of the islets of Langerhans as the precursor, proinsulin, which is processed to form C-peptide and insulin. Both are secreted in equimolar amounts into the portal circulation. The mature insulin molecule comprises two polypeptide chains, the A chain and B chain (21 and 30 amino acids respectively). The two chains are linked together by two inter-chain disulphide bridges. There is also an intra-chain disulphide bridge in the A chain. Secretion of insulin is mainly controlled by plasma glucose concentration, and the hormone has a number of important metabolic actions. Its principal function is to control the uptake and utilisation of glucose in peripheral tissues via the glucose transporter. This and other hypoglycaemic activities, such as the inhibition of hepatic gluconeogenesis and glycogenolysis are counteracted by the hyperglycaemic hormones including glucagon, epinephrine (adrenaline), growth hormone and cortisol.

Insulin concentrations are severely reduced in insulin-dependent diabetes mellitus (IDDM) and some other conditions such as hypopituitarism. Insulin levels are raised in non-insulin-dependent diabetes mellitus (NIDDM), obesity, insulinoma and some endocrine dysfunctions such as Cushing's syndrome and acromegaly.

PRINCIPLE of the test

The Insulin ELISA Kit is a solid phase enzyme-linked immunosorbent assay (ELISA) based on the sandwich principle.

The microtiter wells are coated with a monoclonal antibody directed towards a unique antigenic site on the Insulin molecule.

An aliquot of patient sample containing endogenous Insulin is incubated in the coated well with enzyme conjugate, which is an anti-Insulin antibody conjugated with Biotin. After incubation the unbound conjugate is washed off.

During the second incubation step Streptavidin Peroxidase Enzyme Complex binds to the biotin-anti-Insulin antibody. The amount of bound HRP complex is proportional to the concentration of Insulin in the sample. Having added the substrate solution, the intensity of colour developed is proportional to the concentration of Insulin in the patient sample.

Precautions

- This kit is for in vitro diagnostic use only.
- For information on hazardous substances included in the kit please refer to Material Safety Data Sheets.
- All reagents of this test kit which contain human serum or plasma have been tested and confirmed negative for HIV I/II, HBsAg and HCV by FDA approved procedures. All reagents, however, should be treated as potential biohazards in use and for disposal.
- Avoid contact with Stop Solution containing 0.5 M H_2SO_4 . It may cause skin irritation and burns.
- Never pipet by mouth and avoid contact of reagents and specimens with skin and mucous membranes.
- Do not smoke, eat, drink or apply cosmetics in areas where specimens or kit reagents are handled.
- Wear disposable latex gloves when handling specimens and reagents. Microbial contamination of reagents or specimens may give false results.
- Handling should be in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
- Do not use reagents beyond expiry date as shown on the kit labels.
- All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes.
- Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
- Chemicals and prepared or used reagents have to be treated as hazardous waste according the national biohazard safety guideline or regulation.
- Safety Data Sheets for this product are available upon request directly from the manufacture.
- The Safety Data Sheets fit the demands of: EU-Guideline 91/155 EC.

Kit Components

Contents of the Kit

96

ME E-0931 Microtiter wells

12x8 (break apart) strips, 96 wells; Wells coated with anti-Insulin monoclonal antibody

Standards

	Cat. no.	Standard	Concentration	Volume/Vial
STANDARD A	ME E-0901	Zero Standard	0 µIU/ml	3 ml
STANDARD B	ME E-0902	Standard 1	6.25 µIU/ml	1 ml
STANDARD C	ME E-0903	Standard 2	12.5 µIU/ml	1 ml
STANDARD D	ME E-0904	Standard 3	25 µIU/ml	1 ml
STANDARD E	ME E-0905	Standard 4	50 µIU/ml	1 ml
STANDARD F	ME E-0906	Standard 5	100 µIU/ml	1 ml

The standards are calibrated against international WHO approved Reference material NIBSC 66/304.

CONJUGATE

ME E-0940 Enzyme Conjugate

1 vial, 5 ml, ready to use, mouse monoclonal anti-Insulin conjugated to biotin

ENZYME

ME E-0915 Enzyme Complex

1 vial, 7 ml, ready to use, Streptavidin HRP Complex

SUBSTRATE

FR E-0055 Substrate Solution

1 vial, 14 ml, ready to use, TMB

STOP-SOLN

FR E-0080 Stop Solution

1 vial, 14 ml, ready to use, contains 0.5M H₂SO₄,
Avoid contact with the stop solution. It may cause skin irritations and burns.

WASH-CONC 40x

FR E-0030 Wash Solution

1 vial, 30 ml (40X concentrated), see „Preparation of Reagents“

Note: Additional Zero Standard for sample dilution is available on request.

Equipment and material required but not provided

- A microtiter plate calibrated reader (450±10 nm)
- Calibrated variable precision micropipettes.
- Absorbent paper.
- Aqua dest.

Storage and stability of the Kit

When stored at 2-8°C unopened reagents will retain reactivity until expiration date. Do not use reagents beyond this date.

All opened reagents must be stored at 2-8°C. Microtiter wells must be stored at 2-8°C. Once the foil bag has been opened, care should be taken to close it tightly again.

Preparation of Reagents

Allow all reagents and required number of strips to reach room temperature prior to use.

Wash Solution

Dilute 30 ml of concentrated Wash Solution with 1170 ml deionized water to a final volume of 1200 ml. The diluted Wash Solution is stable for 2 weeks at room temperature.

Disposal of the Kit

The disposal of the kit must be made according to the national official regulations. Special information for this product is given in the Material Safety Data Sheets (see chapter 13).

Damaged Test Kits

In case of any severe damage of the test kit or components, the manufacture have to be informed written, latest one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.

SPECIMEN

Serum or plasma (only Heparin- or citrate plasma) can be used in this assay.
Do not use haemolytic, icteric or lipemic specimens.

Specimen Collection

Serum:

Collect blood by venipuncture (e.g. Sarstedt Monovette # 02.1388.001), allow to clot, and separate serum by centrifugation at room temperature.

Plasma:

Whole blood should be collected into centrifuge tubes containing anti coagulant and centrifuged immediately after collection.

(E.g. for Heparin plasma Sarstedt Monovette – orange cap - # 02.165.001; for Citrate plasma Sarstedt Monovette – green cap - # 02.167.001.)

Specimen Storage

Specimens should be capped and may be stored for up to 5 days at 2-8°C prior to assaying.

Specimens held for a longer time should be frozen only once at -20°C prior to assay. Thawed samples should be inverted several times prior to testing.

Specimen Dilution

If in an initial assay, a specimen is found to contain more than the highest standard, the specimens can be diluted 10-fold or 100 fold with Zero Standard and reassayed as described in Assay Procedure.

For the calculation of the concentrations this dilution factor has to be taken into account.

Example:

- a) dilution 1:10: 10 µl Serum + 90 µl Standard Zero Standard (mix thoroughly)
- b) dilution 1:100: 10 µl dilution a) 1:10 + 90 µl Zero (mix thoroughly).

Test procedure

General Remarks

- All reagents and specimens must be allowed to come to room temperature before use. All reagents must be mixed without foaming.
- Once the test has been started, all steps should be completed without interruption.
- Use new disposal plastic pipette tips for each standard, control or sample in order to avoid cross contamination
- Absorbance is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.

Assay Procedure

1.	Secure the desired number of Microtiter wells in the holder.
2.	Dispense 25 µl of each Standard, controls and samples with new disposable tips into appropriate wells.
3.	Dispense 25 µl Enzyme Conjugate into each well.
4.	Thoroughly mix for 10 seconds. It is important to have a complete mixing in this step.
5.	Incubate for 30 minutes at room temperature without covering the plate.
6.	Briskly shake out the contents of the wells. Rinse the wells 3 times with diluted Wash Solution (400 µl per well). Strike the wells sharply on absorbent paper to remove residual droplets. Important note: The sensitivity and precision of this assay is markedly influenced by the correct performance of the washing procedure!
7.	Add 50 µl of Enzyme Complex to each well.
8.	Incubate for 30 minutes at room temperature.
9.	Briskly shake out the contents of the wells. Rinse the wells 3 times with diluted Wash Solution (400 µl per well). Strike the wells sharply on absorbent paper to remove residual droplets
10.	Add 50 µl of Substrate Solution to each well.
11.	Incubate for 15 minutes at room temperature.
12.	Stop the enzymatic reaction by adding 50 µl of Stop Solution to each well.
13.	Read the OD at 450±10 nm with a microtiter plate reader within 10 minutes after adding the Stop Solution.

Calculation of Results

1. Calculate the average absorbance values for each set of standards, controls and patient samples.
2. Construct a standard curve by plotting the mean absorbance obtained from each standard against its concentration with absorbance value on the vertical(Y) axis and concentration on the horizontal (X) axis.
3. Using the mean absorbance value for each sample determine the corresponding concentration from the standard curve.
4. Automated method: Computer programs using cubic spline, 4 PL (4 Parameter Logistics) or Logit-Log can generally give a good fit.
5. The concentration of the samples can be read directly from this standard curve. Samples with concentrations higher than that of the highest standard have to be further diluted. For the calculation of the concentrations this dilution factor has to be taken into account.

Below is listed a typical example of a standard curve with the Insulin ELISA.

Standard	Optical Units (450 nm)
Standard 0 (0 µIU/ml)	0.03
Standard 1 (6.25 µIU/ml)	0.07
Standard 2 (12.5 µIU/ml)	0.14
Standard 3 (25 µIU/ml)	0.35
Standard 4 (50 µIU/ml)	0.88
Standard 5 (100 µIU/ml)	2.05

Expected values

It is strongly recommended that each laboratory should determine its own normal and abnormal values.

In a study conducted with apparently normal healthy adults, using the Insulin ELISA the following values are observed: 2 to 25 µIU/ml.

Assay Characteristics

Assay Dynamic Range

The range of the assay is between 0 – 100 μ IU/ml.

Specificity of Antibodies (Cross Reactivity)

The cross reactivities were determined by addition of different analytes to serum containing 4 ng/ml ($\approx$ 100 μ IU/ml) Insulin and measuring the apparent Insulin concentration.

Added analyte to a high value serum (4 ng/ml)	Observed Insulin value (ng/ml)	Cross reaction (%)
Porcine Insulin 8 ng/ml	17	> 100
Bovine Insulin 8 ng/ml	17.8	> 100
Dog Insulin 16 ng/ml	17.2	82
Rabbit Insulin 16 ng/ml	14.1	63
Rat Insulin 16 ng/ml	4.0	0
Human Proinsulin 32 ng/ml	4.1	0
Porcine Proinsulin 16 ng/ml	4.0	0
Bovine Proinsulin 16 ng/ml	4.1	0

Analytical Sensitivity

The analytical sensitivity was calculated from the mean plus two standard deviations of twenty (20) replicate analyses of Zero Standard and was found to be 1.76 μ IU/ml.

Precision

Intra Assay Variation

The within assay variability is shown below:

Sample	n	Mean (μ IU/ml)	CV (%)
1	20	17.5	2.6
2	20	66.4	1.8

Inter Assay Variation

The between assay variability is shown below:

Sample	n	Mean (μ IU/ml)	CV (%)
1	12	17.4	2.9
2	12	66.9	6.0

Accuracy

Quality Control

It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results. Use controls at both normal and pathological levels. The controls and the corresponding results of the QC-Laboratory are stated in the QC certificate added to the kit. The values and ranges stated on the QC sheet always refer to the current kit lot and should be used for direct comparison of the results.

It is also recommended to make use of national or international Quality Assessment programs in order to ensure the accuracy of the results.

Employ appropriate statistical methods for analysing control values and trends. If the results of the assay do not fit to the established acceptable ranges of control materials patient results should be considered invalid.

In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods.

After checking the above mentioned items without finding any error contact your distributor or the manufacture directly.

Recovery

Samples have been spiked by adding Insulin solutions with known concentrations in a 1:1 ratio. The expected values were calculated by addition of half of the values determined for the undiluted samples and half of the values of the known solutions. The % Recovery has been calculated by multiplication of the ratio of the measurements and the expected values with 100.

Sample	Added Concentration 1:1 (v/v) (μIU/ml)	Measured Conc. (μIU/ml)	Expected Conc. (μIU/ml)	Recovery (%)
1	100	21.24		
	50	66.41	60.62	109.6
	25	38.80	35.62	108.9
	12.5	23.38	23.12	101.1
2	100	17.37	16.87	102.9
	50	69.04		
	25	84.58	84.52	100.1
	12.5	58.39	59.52	98.1
	100	43.17	47.02	91.8
	50	37.48	40.77	91.9
	25			
	12.5			

Linearity

Sample	Dilution	Measured Conc. (μIU/ml)	Expected Conc. (μIU/ml)	Recovery (%)
1	None	21.24	21.24	
	1:2	9.40	10.62	88.5
	1:4	5.23	5.31	98.5
	1:8	2.81	2.66	105.9
	1:16	1.47	1.33	110.3
2	None	69.04	69.04	
	1:2	30.51	34.52	88.4
	1:4	17.60	17.26	102.0
	1:8	8.73	8.63	101.2
	1:16	4.76	4.32	110.4

Limitations of Use

Interfering Substances

Any improper handling of samples or modification of this test might influence the results. Haemoglobin (up to 4 mg/ml), Bilirubin (up to 0.5 mg/ml) and Triglyceride (up to 30 mg/ml) have no influence on the assay results.

Drug Interferences

Until today no substances (drugs) are known to us, which have an influence to the measurement of Insulin in a sample.

High-Dose-Hook Effect

No hook effect was observed in this test up to 1600 μIU/ml of Insulin.

Legal Aspects

Reliability of Results

The test must be performed exactly as per the manufacturer's instructions for use. Moreover the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable national standards and/or laws. This is especially relevant for the use of control reagents. It is important to always include, within the test procedure, a sufficient number of controls for validating the accuracy and precision of the test. The test results are valid only if all controls are within the specified ranges and if all other test parameters are also within the given assay specifications. In case of any doubt or concern please contact the manufacture.

Therapeutical Consequences

Therapeutical consequences should never be based on laboratory results alone even if all test results are in agreement with the items as stated under point 10.1. Any laboratory result is only a part of the total clinical picture of a patient.

Only in cases where the laboratory results are in acceptable agreement with the overall clinical picture of the patient should therapeutical consequences be derived.

The test result itself should never be the sole determinant for deriving any therapeutical consequences.

Liability

Any modification of the test kit and/or exchange or mixture of any components of different lots from one test kit to another could negatively affect the intended results and validity of the overall test. Such modification and/or exchanges invalidate any claim for replacement.

Claims submitted due to customer misinterpretation of laboratory results subject to point 10.2. are also invalid. Regardless, in the event of any claim, the manufacturer's liability is not to exceed the value of the test kit. Any damage caused to the test kit during transportation is not subject to the liability of the manufacturer.

REFERENCES

1. Flier, J. S., Kahn, C. R. and Roth, J. (1979). Receptors, antireceptor antibodies and mechanisms of insulin resistance; N. Engl. J. Med., 300, 8, 413-419.
2. Frier, B. M., Ashby, J. P., Nairn, I. M. and Bairs, J.D. (1981). Plasma insulin, C-peptide and glucagon concentrations in patients with insulin-independent diabetes treated with chlorpropamide. Diab. metab., 7,1, 45-49.
3. Judzewitsch, R. G., Pfeifer, M. A., Best, J. D., Beard J. C., Halter, J. B. and Porte D. Jr. (1982). Chronic Chlorpropamide therapy of noninsulin-dependent diabetes augments basal and stimulated insulin secretion by increasing islet sensitivity to glucose. J. Clin. End. and Metab. 55, 2, 321-328.
4. Kosaka, K., Hagura, R. and Kuzuya, T. (1977). Insulin responses in equivocal and definite diabetes, with special reference to subjects who had mild glucose intolerance but later developed definite diabetes. Diabetes 26, 10, 944-952.
5. Starr, J. Il, Mako, M. E., Juhn, D. and Rubenstein, A. H. (1978). Measurement of serum proinsulin-like material: cross-reactivity of porcine and human proinsulin in the insulin radioimmunoassay, J. Lab. Clin. Med. 91,4, 691-692.

Symbols:

	Storage temperature		Manufacturer		Contains sufficient for <n> tests
	Expiry date		Batch code		For in-vitro diagnostic use only!
	Consult instructions for use		Content		CE labelled
	Caution		Catalogue number		For research use only!